

試験結果(Test Data)

**SAMPLE**

型名 (Model No.) : 6243  
 製造番号 (Serial No.) : 123456789

1. 電圧発生 [4 1/2桁表示] (Voltage Source, 4 1/2 digit display)

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
320 mV	300.00 mV	299.71 mV	300.00 mV	300.29 mV	良(Pass)
	-300.00 mV	-300.29 mV	-300.00 mV	-299.71 mV	良(Pass)
3.2 V	3.0000 V	2.9985 V	3.0000 V	3.0015 V	良(Pass)
	-3.0000 V	-3.0015 V	-3.0000 V	-2.9985 V	良(Pass)
32 V	30.000 V	29.985 V	30.000 V	30.015 V	良(Pass)
	-30.000 V	-30.015 V	-30.000 V	-29.985 V	良(Pass)
110 V	100.00 V	99.94 V	100.00 V	100.06 V	良(Pass)
	-100.00 V	-100.06 V	-100.00 V	-99.94 V	良(Pass)

2. 電流発生 [4 1/2桁表示] (Current Source, 4 1/2 digit display)

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
32 μA	30.000 μA	29.981 μA	29.999 μA	30.019 μA	良(Pass)
	-30.000 μA	-30.019 μA	-30.001 μA	-29.981 μA	良(Pass)
320 μA	300.00 μA	299.81 μA	300.00 μA	300.19 μA	良(Pass)
	-300.00 μA	-300.19 μA	-300.01 μA	-299.81 μA	良(Pass)
3.2 mA	3.0000 mA	2.9981 mA	3.0000 mA	3.0019 mA	良(Pass)
	-3.0000 mA	-3.0019 mA	-3.0001 mA	-2.9981 mA	良(Pass)
32 mA	30.000 mA	29.981 mA	29.999 mA	30.019 mA	良(Pass)
	-30.000 mA	-30.019 mA	-30.001 mA	-29.981 mA	良(Pass)
320 mA	300.00 mA	299.75 mA	300.00 mA	300.25 mA	良(Pass)
	-300.00 mA	-300.25 mA	-300.01 mA	-299.75 mA	良(Pass)
2 A	1000.0 mA	998.4 mA	999.9 mA	1001.6 mA	良(Pass)
	-1000.0 mA	-1001.6 mA	-1000.2 mA	-998.4 mA	良(Pass)

3. 電圧測定 [5 1/2桁表示] (Voltage Measurement, 5 1/2 digit display)

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
320 mV	300.000 mV	299.810 mV	299.998 mV	300.190 mV	良(Pass)
	-300.000 mV	-300.190 mV	-300.007 mV	-299.810 mV	良(Pass)
3.2 V	3.00000 V	2.99895 V	2.99999 V	3.00105 V	良(Pass)
	-3.00000 V	-3.00105 V	-2.99999 V	-2.99895 V	良(Pass)
32 V	30.0000 V	29.9900 V	30.0000 V	30.0100 V	良(Pass)
	-30.0000 V	-30.0100 V	-29.9999 V	-29.9900 V	良(Pass)
110 V	100.000 V	99.962 V	100.000 V	100.038 V	良(Pass)
	-100.000 V	-100.038 V	-99.999 V	-99.962 V	良(Pass)

4. 電流測定 [5 1/2桁表示] (Current Measurement, 5 1/2 digit display)

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
32 μA	30.0000 μA	29.9830 μA	30.0010 μA	30.0170 μA	良(Pass)
	-30.0000 μA	-30.0170 μA	-29.9993 μA	-29.9830 μA	良(Pass)
320 μA	300.000 μA	299.830 μA	300.007 μA	300.170 μA	良(Pass)
	-300.000 μA	-300.170 μA	-299.991 μA	-299.830 μA	良(Pass)
3.2 mA	3.00000 mA	2.99830 mA	3.00007 mA	3.00170 mA	良(Pass)
	-3.00000 mA	-3.00170 mA	-2.99992 mA	-2.99830 mA	良(Pass)
32 mA	30.0000 mA	29.9830 mA	30.0007 mA	30.0170 mA	良(Pass)
	-30.0000 mA	-30.0170 mA	-29.9993 mA	-29.9830 mA	良(Pass)
320 mA	300.000 mA	299.770 mA	300.006 mA	300.230 mA	良(Pass)
	-300.000 mA	-300.230 mA	-299.989 mA	-299.770 mA	良(Pass)
2 A	1000.00 mA	998.60 mA	1000.09 mA	1001.40 mA	良(Pass)
	-1000.00 mA	-1001.40 mA	-999.89 mA	-998.60 mA	良(Pass)

※ 規格(上限・下限)は、温度23℃±5度、相対湿度85%以下における、1年間の保証確度の値です。  
 Accuracy(Min/Max) is guaranteed for the duration of one year under the condition with temperature of 23℃±5℃, humidity of less than 85%.