

試験結果
Test Data

SAMPLE

管理番号: 0123456789
(Control No.)

型名 (Model No.) 7 3 5 1 E
製造番号 (Serial No.) 1 2 3 4 5 6 7 8 9

1、直流電圧測定 (DC Voltage Measurement) (5 1/2 digits display, IT:100ms(SLOW1))

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
200 mV	190.000 mV	189.971 mV	189.998 mV	190.029 mV	良 (Pass)
	-190.000 mV	-190.029 mV	-189.994 mV	-189.971 mV	良 (Pass)
2000 mV	1900.00 mV	1899.77 mV	1900.00 mV	1900.23 mV	良 (Pass)
	-1900.00 mV	-1900.23 mV	-1899.98 mV	-1899.77 mV	良 (Pass)
20 V	19.0000 V	18.9967 V	18.9993 V	19.0034 V	良 (Pass)
	-19.0000 V	-19.0034 V	-18.9993 V	-18.9967 V	良 (Pass)
200 V	190.000 V	189.967 V	189.997 V	190.034 V	良 (Pass)
	-190.000 V	-190.034 V	-189.996 V	-189.967 V	良 (Pass)
1000 V	1000.00 V	999.80 V	1000.00 V	1000.20 V	良 (Pass)
	-1000.00 V	-1000.20 V	-1000.00 V	-999.80 V	良 (Pass)

2、交流電圧測定 (AC Voltage Measurement) (AC Coupling, 5 1/2 digits display, IT:100ms(SLOW1), Frequency:1kHz)

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
200 mV	190.000 mV	189.710 mV	190.018 mV	190.290 mV	良 (Pass)
2000 mV	1900.00 mV	1897.10 mV	1899.98 mV	1902.90 mV	良 (Pass)
20 V	19.0000 V	18.9710 V	19.0005 V	19.0290 V	良 (Pass)
200 V	190.000 V	189.710 V	190.011 V	190.290 V	良 (Pass)
700 V	700.00 V	698.30 V	700.04 V	701.70 V	良 (Pass)

3、直流電流測定 (DC Current Measurement) (5 1/2 digits display, IT:100ms(SLOW1))

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
200 mA	190.000 mA	189.933 mA	189.986 mA	190.067 mA	良 (Pass)
	-190.000 mA	-190.067 mA	-189.989 mA	-189.933 mA	良 (Pass)
2000 mA	1900.00 mA	1898.99 mA	1899.95 mA	1901.01 mA	良 (Pass)
	-1900.00 mA	-1901.01 mA	-1899.88 mA	-1898.99 mA	良 (Pass)
10 A	10.0000 A	9.9844 A	9.9992 A	10.0156 A	良 (Pass)
	-10.0000 A	-10.0156 A	-9.9995 A	-9.9844 A	良 (Pass)

4、交流電流測定 (AC Current Measurement) (AC Coupling, 5 1/2 digits display, IT:100ms(SLOW1), Frequency:1kHz)

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
200 mA	190.000 mA	189.330 mA	189.982 mA	190.670 mA	良 (Pass)
2000 mA	1900.00 mA	1892.35 mA	1899.78 mA	1907.65 mA	良 (Pass)
10 A	10.0000 A	9.9450 A	10.0015 A	10.0550 A	良 (Pass)

5、抵抗測定 (Resistance Measurement) (2-wire, 5 1/2 digits display, IT:100ms(SLOW1))

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
200 Ω	190.000 Ω	189.954 Ω	190.049 Ω	190.046 Ω	否 (Fail)
2000 Ω	1900.00 Ω	1899.70 Ω	1899.97 Ω	1900.30 Ω	良 (Pass)
20 kΩ	19.0000 kΩ	18.9970 kΩ	18.9982 kΩ	19.0030 kΩ	良 (Pass)
200 kΩ	190.000 kΩ	189.957 kΩ	189.981 kΩ	190.043 kΩ	良 (Pass)
2000 kΩ	1900.00 kΩ	1899.33 kΩ	1899.82 kΩ	1900.67 kΩ	良 (Pass)
20 MΩ	19.0000 MΩ	18.9610 MΩ	18.9981 MΩ	19.0390 MΩ	良 (Pass)
200 MΩ	100.000 MΩ	98.490 MΩ	99.983 MΩ	101.510 MΩ	良 (Pass)

6、周波数測定 (Frequency Measurement) (5 1/2 digits display, Gate Time:1000ms)

試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
199.900 kHz	199.860 kHz	199.913 kHz	199.940 kHz	良 (Pass)